

BeneHeart C1A

Automated External Defibrillator (AED)



Defibrillator

Operations	Semi-automatic and fully automatic versions
Waveform	Biphasic Truncated exponential (BTe), with automated voltage and duration compensation for patient impedance
Range of selected energy	100 to 360 J (adult) 10 to 100 J (child)
Energy default	200-300-360 J (adult) 50-70-100 J (child) Default configuration meets AHA/ERC Guidelines 2015.
Energy accuracy	± 2 J or ± 10 % of setting, whichever is greater
Power on time	< 2 seconds
ECG analysis time	< 5 seconds
Charge time	0 seconds (as device is pre-charged during ECG analysis)
Time from power on to shock ready	< 8 seconds (200J, new battery, $20 \pm 5^\circ\text{C}$)
Mindray shockable rhythm analysis algorithm	Acquires and analyzes the patient's ECG signals to determine whether or not to give a defibrillation shock
Sensitivity and specificity	Meets AAMI DF80 specifications and IEC 60601-2-4 specifications
Patient impedance range	25 to 300 Ω

User Prompts

User prompts	Voice prompts
CPR coaching	Voice guide CPR metronome CPR real-time feedback ¹
CPR protocol	Meets AHA/ERC Guidelines 2015 and/or can be configured locally

Controls

Lid release/ON-OFF	Controls device power on/off
Shock button	Delivers energy when button presses by the user (semi-automatic only)
Adult/child mode switch	Switch to child mode for reduced energy and appropriate CPR guidance

Language button

Optional feature allows the user to switch between max. 3 languages

Physical Characteristics

Dimension	210 mm (w) x 286 mm (d) x 78 mm (h)
Weight	2.0 kg (including one battery)

Environmental

Dust/water resistance	IP55
Temperature	Operating: -5 to 50°C Short-term storage: -30 to 70°C for a maximum of 7 days Long-term storage: 15 to 35°C
Humidity	Operating/storage: 5 to 95 % (non-condensing)
Altitude	Operating/storage: -381 m to $+4575$ m
Shock	RTCA-DO-160G-2010, Section 7 IEC60601-1-12, 10.1.3, 10.1.4
Vibration	MIL-STD-810G-2008, method 514.6, Category 13, Category 14, Category 20, Category 24 EN13718-1, 4.7.2
Bump	EN1789, 6.3.4.2 EN13718-1, 4.7.2
Drop	1.5 m
EMC	IEC60601-1-2: 2014 EN13718-1, 4.5.7 IEC 60601-1-12, 11

Battery

Type	Lithium manganese dioxide (Li/MnO ₂), disposable, 4200 mAh
Standby life	6 years (at $20 \pm 5^\circ\text{C}$, performing auto test every week, not in use, not sending self-test report) 5 years (at $20 \pm 5^\circ\text{C}$, performing auto test every day, not in use, not sending self-test report)
Capacity	With new battery at $20 \pm 5^\circ\text{C}$: ≥ 15 hours of operating times; provides max. 400 shocks @200J (± 3 shocks < 1 minute)
Replace battery	Min. 10 shocks at 200 J and 30 minutes



indication	of operating time (at 20±5 °C, typical).
Weight	300 g
Electrode Pads	
Type	Pre-connected, disposable, for adult/child
Shelf life	5 years (from date of manufacture)
CPR Sensor ²	
Weight	Approximately 180 g (without battery)
Thickness	17.5 to 19 mm
Automatic Self-test	
Auto-test	Daily, weekly, monthly, quarterly

P/N:ENG-C2 Datasheet -210285X2P-20190830
©2019 Shenzhen Mindray Bio-Medical Electronics Co.,Ltd. All rights reserved.

¹ Requiring to configure with CPR sensor

² For further information about the availability of CPR sensor, please contact with your local sales representatives.

³ For further information about the availability of 4G data transfer and AED-Alert™ 2.0 system, please contact with your local sales representatives.

www.mindray.com

P/N:ENG-C2 Datasheet -210285X2P-20190830
©2019 Shenzhen Mindray Bio-Medical Electronics Co.,Ltd. All rights reserved.

<https://defib4life.co.uk>
michelle@defib4life.co.uk
07919 440 044

